

@ivoclarvivadent.com>

@aefu.ch

25.8.2014

WG: Ask Us a Question

Dear Mr Fortner

Thank you for your inquiry.

I forwarded your question to our R&D department and got following reply.

- 0.3 % triclosan is used as stabilizer in Telio CS Inlay
- Actually it is not planned to discontinue the formulation

Hope this will answer your questions fully.

Best regards

Renate Demmel **Ivoclar Vivadent AG** Assistant Professional Services

Bendererstrasse 2 9494 Schaan - Liechtenstein

www.ivoclarvivadent.com